

# Declaration of Conformity

(according to ISO/IEC Guide 22 and EN 45014)

**Manufacturer's Name:** Medical Graphics Corporation

**Manufacturer's Address:** 350 Oak Grove Parkway  
Saint Paul, Minnesota 55127 USA

*Declares under our sole responsibility that the following product:*

**Product Name:** VO2000 Metabolic Analysis System

**Model Number(s):** 80080X-00X

**Model Option(s):** Ambulatory without Telemetry, Ambulatory with Telemetry

**Manufactured** 9-15-2011 **and beyond**

*to which this declaration relates, complies with European Council Directive 93/42/EEC Medical Device Directive Annex II and is in conformity with the following standard(s) or other normative document(s):*

ISO 13485

EN 60601-1 Medical Electrical Equipment, General Requirements for Safety

EN 60601-1-2 General Requirements for Safety, Electromagnetic  
Compatibility Requirements and Tests


*and is classified Iia per Annex IX, rule 10 following the provisions of 93/42/EEC, European Medical Device Directive.*

**E.C. Certificate No. CE 94534**

Notified Body: BSI Product Service  
Maylands Avenue  
Hemel Hempstead, Hertfordshire  
United Kingdom


**CE 0086**

**Date** 9/15/11



**Gregg O. Lehman**  
President and C.E.O.

**Medical Graphics Corporation**



**MEDGRAPHICS™**  
Cardiorespiratory Diagnostics